


Malaria Generic Pf-Pan RDT Results Guide



NEGATIVE RESULT



Negative

POSITIVE RESULTS



P. falciparum



Non-falciparum (*P. vivax*, *P. ovale*,
P. malariae or a mixed infection of these)



P. falciparum



P. falciparum monoinfection
or 'Mixed' Infection

NOTE: Test is positive even if the line in the test window is faint.

INVALID RESULTS



No Control Line



No Control Line

Repeat the test using a new RDT if no control line appears.